

Section Subject	UTMB On-line Documentation Healthcare Epidemiology Policies and Procedures	01.11-Policy
Topic	01.11 - Evacuation of Laser Plume	08.31.15-Revised 1997-Author

01.11 Evacuation of Laser Plume

Purpose To provide effective evacuation of laser plume and smoke from electrosurgical units. According to several studies, laser plume (smoke) is potentially hazardous and should be effectively evacuated (1-5).

Audience All healthcare personnel who use the laser or electrosurgical unit

Guideline During procedures using a laser or electrosurgical unit, the thermal destruction of tissue creates a smoke byproduct. Large smoke plumes can be captured and disposed of via smoke evacuators with an in-line filter. Healthcare workers should wear protective devices when using a laser or electrosurgical unit.

Procedure, Smoke Evacuation System in the OR

- The smoke evacuation system must be adequate to handle the amount of plume produced during surgical procedures. In-line suction filters may be used for small amounts of plume (e.g., for microlaryngoscopic vaporization of vocal cord polyps). A smoke evacuation system with an evacuation hose will be used for large amounts of plume. Endoscopic/laparoscopic smoke evacuation requires special efforts.
- The smoke evacuator should be ON (activated at all times when airborne particles are produced during all laser and electrosurgical procedures).
- The smoke evacuator or room suction hose nozzle inlet must be kept within 1 centimeter of the surgical site to effectively capture airborne contaminants generated by surgical devices (2, 4).
 - In-Line suction - An in-line filter is placed between the suction canister and the wall connection.
 - An in-line suction smoke evacuation filter can become blocked by the particulate matter in large amounts of plume. The filter should be monitored and changed as needed.
 - Work practices - Hold the smoke evacuation suction tube close (within 1 centimeter) to the tissue interaction site to remove as much plume as possible. Surgical plume contains extremely small particulate matter and may contain viable cells.
 - When purge gas flow is used with the CO2 laser or a fiber delivery device, the smoke evacuation tube must constantly be held close to the laser-tissue interaction site because the gas flow will tend to spread the plume.
 - If the smoke evacuation foot pedal is available, the assistant can operate it.

Section Subject	UTMB On-line Documentation Healthcare Epidemiology Policies and Procedures	01.11-Policy
Topic	01.11 - Evacuation of Laser Plume	08.31.15-Revised 1997-Author

- Smoke evacuation hook up - Ensure the proper operation of the smoke evacuator prior to the beginning of the case. Check the plume filter, and if needed, install a clean filter. For the maximum filter life, operate the unit at the lowest settings that will adequately vacuum away the smoke plume. As the filter begins to clog, increasing the suction level will permit the filter to be used for a longer period of time. Filters should be changed as recommended by the manufacturer. At the end of the procedure, the filter is considered infectious waste and should be disposed of properly. Standard Precautions should be used when changing filters.
- Endoscopic/Laparoscopic Smoke Evacuators - Instruments such as suction tubes, help decrease the retention of plume inside a body cavity or organ. A low-pressure suction valve can be used to gently remove plume during a laparoscopic procedure without significantly reducing the pneumoperitoneum.

Laser Use

Use in general medical/ surgical areas and clinics

- The smoke evacuation systems should be used to evacuate large volumes of smoke. Routine use of the laser or electro-surgical unit in a medical or surgical inpatient area or in a clinic requires only the use of masks and eye protection.

Masks and Protection of Eyes

- Masks, with filtering capacity of particulate matter between 2 and 5 microns, should be worn by healthcare workers during any nonendoscopic procedure that generates laser plume or smoke from an electro-surgical procedure.
- The sides of the mask should conform to the face adequately. The mask should be tight fitting so that there is no leakage of air around the edges. Standard Precautions should be followed for the removal/disposal of masks.
- Goggles or glasses with side shields should be worn by members of the laser team (6,7).

Decontamination

Decontamination of the treatment room after the procedure

- All patients are considered potentially infected with bloodborne pathogens. All procedures, therefore, are considered contaminated, and the same environmental cleaning protocols should be implemented for all procedures.

Section Subject	UTMB On-line Documentation Healthcare Epidemiology Policies and Procedures	01.11-Policy
Topic	01.11 - Evacuation of Laser Plume	08.31.15-Revised 1997-Author

- References
1. Garden, J.M., O'Banion, M.K., Shelnitz, L.S., et al. "Papillomavirus in the vapor of carbon dioxide laser-treated verrucae," *JAMA* (1988) 259, 1199-1202.
 2. Sawchuck, W.S., Weber, P.J., Lowy, D.R., Dzubou, L.M. "Infectious papillomavirus in the vapor of warts treated with carbon dioxide laser or electrocoagulation: detection and protection," *Journal of the American Academy of Dermatology* (1989) 21, 41-49.
 3. Andre, P., Orth, G., Evenon, P., Guillaume, J.C., Avril, M.F. "Risk of papillomavirus infection in carbon dioxide laser treatment of genital lesions," *Journal of the American Academy of Dermatology* (1990) 22, 131-132.
 4. Fereczy, A., Bergeron, C., Richart, R.M. "Human papillomavirus DNA in CO2 laser generated plume of smoke and its consequences to the surgeon," *Obstetrics and Gynecology* (1990) 75, 114-118.
 5. Kashima, H.K., Kessis, T., Mounts, P., Shaw, K. "Polymerase chain reaction identification of human papillomavirus DNA in CO2 laser plume from recurrent respiratory papillomatosis," *Otolaryngology Head and Neck Surgery* (1991) 104, 191-195.
 6. Bergbrant, I., Samuelsson, L., Olosson, S., Jonassen, F., Ricksten, A. "Polymerase chain reaction for monitoring human papillomavirus contamination of medical personnel during treatment of genital warts with CO2 laser and electrocoagulation," *Acta Dermatol Venerologica* (Stockholm) (1994) 74, 393-395.
 7. Abramson, A.L., DiLorenzo, T.P., Steinberg, B.M. "Is papillomavirus detectable in the plume of laser-treated laryngeal papilloma?" *Archives of Otolaryngology Head and Neck Surgery* (1990) 116, 604-607.
 8. Ball, K.A. (1995). *Laser: The perioperative challenge* (2nd ed.). St. Louise, MO: C.V. Mosby.
 9. American National Standards Institute, "*The safe use of laser in health care facilities*," Z136.3. New York: American National Standards Institute, 1996.
 10. National Institute of Occupational Safety and Health, *Control of Smoke from Laser/Electric Surgical Procedures*. Cincinnati: OSHA [NIOSH] publ no 95-128, 1996.
 11. Lindsey C, Hutchinson M, Mellor G. The nature and hazards of diathermy plumes: a review. *AORN J*. 2015;101:428-442.